



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/720,656	11/24/2003	George Frey	MSDI-247/PC845.00	4638
52196	7590	08/22/2007		
KRIEG DEVAULT LLP ONE INDIANA SQUARE, SUITE 2800 INDIANAPOLIS, IN 46204-2709			EXAMINER SCHILLINGER, ANN M	
			ART UNIT 3738	PAPER NUMBER
			MAIL DATE 08/22/2007	DELIVERY MODE PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

H

Office Action Summary	Application No. 10/720,656	Applicant(s) FREY ET AL.	
	Examiner Ann Schillinger	Art Unit 3738	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 29 May 2007.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-48 and 57-64 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-48 and 57-64 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. _____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application. |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1-4, 7, 15, and 16 are rejected under 35 U.S.C. 102(b) as being anticipated by Liu et al. (US Pat. No. 6,083,154). Liu et al. discloses the following of claim 1: a surgical instrumentation system to provide a surgical approach to a patient's spine, comprising: a frame (8) including a first portion (left side) lying in a first plane and a second portion (right side) lying in a second plane, said second plane forming an angle with said first plane; and a number of retractors (2b) attached to a frame, at least one of said retractors being attached to said first portion of said frame and extending transversely to said first plane and at least one other of said retractors being attached to said second portion of said frame and extending transversely to said second plane (see Figure 1), wherein said first and second portions of said frame each include a recess (8a) to receive clamping devices (7, 12, 22) coupled to respective ones of said retractors, said clamping devices each including a foot (21) with a pair of arms (22a, 22b) located on opposite sides of said frame and a receptacle defined between said pair of arms for receiving said frame between said pair of arms (see Figure 1) with said clamping devices being slidable from said respective recess along a respective one of said first and second portions of said frame for attachment to said respective portion of said frame at a selected portion of said frame at a selected position therealong spaced from said recess (col. 4, line 55 through col. 5, line 37).

Art Unit: 3738

Liu et al. discloses the limitations of claim 2 where the undercut portion of the frame is the outward facing side and the outward facing arm has element 23 to prevent pivoting.

Liu et al. discloses the limitations of claims 3, 4, 15, and 16 as shown in Figure 1.

Liu et al. discloses the limitations of claim 7 as shown in Figure 7.

Claims 17-21, 26-40 are rejected under 35 U.S.C. 102(b) as being anticipated by Koros et al. (US Pat. No. 5,944,658). Koros et al. discloses the following of claim 17: a surgical instrumentation system to provide a surgical approach to a patient's spine, comprising: first and second anchors (54, 56) engageable to first and second vertebrae of the spine; a frame (10, 12) lying in at least one plane; a retractor (20, 18) attachable to said frame, said retractor including a blade portion (26, 28) extending transversely to said at least one plane, said blade portion including a tissue contacting surface (lower surfaces) adapted to contact and retract tissue from the surgical approach; a first distractor mechanism attachable to said frame and extending transversely to said at least one plane, said first distractor mechanism (50) including a distal end engageable to said first anchor (see Figure 1) with said first distractor mechanism in pivotal relation to the first vertebra when said first anchor is engaged with the first vertebra (col. 5, line 53 through col. 6, line 7); and a second distractor mechanism (52) attachable to said frame and extending transversely to said at least one plane, said second distractor mechanism including a distal end engageable to said second anchor (see Figure 1) and further comprising at least one adjustment mechanism (48) engaged to at least of said first and second distractor mechanisms, wherein said at least one adjustment mechanism includes a shaft having a distal end pivotally coupled with said at least one of said first and second distractor mechanisms at a pivoting

Art Unit: 3738

coupling location adjacent a proximal end of said at least one distractor mechanisms, said shaft extending away from said pivoting coupling location toward said frame and into clamping device movable along said frame and operable to clampingly engage said adjustment mechanism to said frame (see Figure 1).

Koros et al. discloses the limitations of claims 18-21 where the anchors are screws 54 and 56; and the yokes (58, 60) and socket portions can be seen in Figure 1.

Koros et al. discloses the following of claim 26: a system of claim 22, wherein said adjustment mechanisms each include: an adjustment handle; a shaft assembly extending from said adjustment handle (see Figure 1); and an engagement member (36) at an end of said shaft assembly opposite said adjustment handle.

Koros et al. discloses the limitations of claims 27-40 as shown in Figure 1.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 5-14 are rejected under 35 U.S.C. 103(a) as being unpatentable over Liu et al. in view of Koros et al. (US Pat. No. 5,944,658). Liu et al. discloses the invention substantially as claimed, however, Liu et al. does not disclose the claimed distractor mechanisms and their features. Koros et al. teaches these distractor mechanisms in columns 4 and 5 for the purpose of manipulating the vertebrae. Therefore, it would have been obvious to one of ordinary skill in the

Art Unit: 3738

art at the time the invention was made to include distractor mechanisms in order to manipulate the vertebrae.

Claims 22-25, 41-48, and 57-64 are rejected under 35 U.S.C. 103(a) as being unpatentable over Koros et al. in view of Liu et al. (US Pat. No. 6,083,154). Koros et al. discloses the invention substantially as claimed, however, Koros et al. does not disclose multiple adjustment members for each of the retractor and distractor mechanisms. Liu et al. teaches having these adjustment members in col. 3, lines 7-40, and col. 4, line 55 through col. 5, line 37 for the purpose of placing the devices the desired position. Therefore, it would have been obvious to one of ordinary skill in the art at the time the invention was made to include these adjustment members in order to place the devices the desired position.

Response to Arguments

Applicant's arguments with respect to claims 1-48 and 57-64 have been considered but are moot in view of the new ground(s) of rejection.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Ann Schillinger whose telephone number is (571) 272-6652. The examiner can normally be reached on Mon. thru Fri. 9 a.m. to 4 p.m..

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Corrine McDermott can be reached on (571) 272-4754. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3738

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Ann Schillinger
August 18, 2007


ANH TUAN T. NGUYEN
SUPERVISORY PATENT EXAMINER
